R&S®NRQ6 FREQUENCY SELECTIVE POWER SENSOR



A milestone in power measurements





ROHDE&SCHWARZ

Make ideas real

Product Brochure Version 04.01

AT A GLANCE

The R&S[®]NRQ6 combines the accuracy of a power meter with excellent dynamic range. It performs extremely precise and fast power measurements down to -130 dBm.

The R&S[®]NRQ6 is based on receiver technology and can perform band-limited power measurements – i.e. power measurements on a selected transmission channel – down to –130 dBm. The R&S[®]NRQ6 delivers high-precision, high-speed measurements beyond the limits of currently available power meters.

In addition to conventional continuous average measurements, the R&S®NRQ6 has a trace display function and also performs ACLR measurements – a common mobile communications application. Using the optional R&S®NRQ6-K1 I/Q data interface, I/Q data can be downloaded from the power sensor to a PC for further analysis.

The R&S®NRQ6-K2 power servoing option enables a dedicated high-speed remote control channel for fast power servoing between an R&S®NRQ6 and an R&S®SGT100A RF vector RF source. This significantly improves the typical setting time via SCPI. Using additionally the R&S®NRQ6-K3 phase coherent measurements option, complex phase coherent measurements can be carried out by configuring one master R&S®NRQ6 and one or multiple slave R&S®NRQ6.

The R&S[®]NRQ6 is controlled via LAN, requiring power over Ethernet (PoE+). The sensor's integrated web server makes it possible to operate the GUI without any extra software – all that is needed is a PC with a web browser. The intuitive GUI is well structured and easy to operate thanks to diverse autoset functions.



Key facts

- ► Frequency selective power measurements
- ► Frequency range: 50 MHz to 6 GHz
- ► Power measurement range: -130 dBm to +20 dBm
- Automatic frequency and bandwidth detection
- ▶ 100 MHz measurement bandwidth
- ► Continuous average, trace and ACLR measurements
- ► I/Q data capturing for RF vector signal analysis
- ► Fast power servoing with R&S[®]SGT100A
- Phase coherent measurements for low-power modulated signals

BENEFITS

Diverse measurement functions

► page 4

Easy operation

► page 6

High measurement speed at lowest signal levels ▶ page 9



DIVERSE MEASUREMENT FUNCTIONS

Settings Help Test Execution - Test	tplan Editor	🎾 Results Viewer 🖧 1	festprocedure Editor o System Configurator	ROHDE&SCH	WARZ
st System: System 1	•	Single Run	Continuous Run		Help
st accomplished:		Test S	tep #: 0 Test Step Id: 0 Repetition:	0/0 DUT Count: 0 Report:	
og Viewer					× 0
Q, Search	Clear			0	Help
Drag a column header and drop	it here to group by	that column			10
imestamo	TestStepid ¥	Source Y	Message		-
018-02-21 13:29:15.418	8	QuickStepEngine			
018-02-21 13:29:19.456		QuickStepEngine	Initialization successfully in '5.321' sec		
018-02-21 11:29:19 457					
018-02-21 13:29:19.457	•	Quickstepengine	starting Te	st	
018-02-21 13:29:19.457		Quickstepengine	Starting Execution Phase: 'Testrum Before'		
018-02-21 13:29:25.285		QuickStepEngine	Finished Execution Phase: 'Testrun Before'		
018-02-21 13:29:25.285		QuickStepEngine	New DUT-Loop started: DUT Count 1		
018-02-21 13:29:25.290		QuickstepEngine	Starting TestStep with Id: 1. Test Procedur		
018-02-21 13:29:27.438		QuickStepEngine	Starting TestStep with Id: 1. Test Procedur		
018-02-21 13:29:29.346		QuickStepEngine	Starting TestStep with Id: 1. Test Procedur		
018-02-21 13:29:31.232		QuickStepEngine	Starting TestStep with Id: 1. Test Procedur		
1018-02-21 13:29:33.117 1018-02-21 13:29:35.009		QuickStepEngine QuickStepEngine	Starting TestStep with Id: 1. Test Procedur Starting TestStep with Id: 1. Test Procedur		
018-02-21 13:29:35.009 018-02-21 13:29:36.812		QuickStepEngine	Rep 1 Time 11.522136 S-> Seq count 6 -> Time		
018-02-21 13:29:36.812		QuickStepEngine	Reptitions completed!	e per lesisiep 1.9204 5 -> blockrunct	101 000
018-02-21 13:29:36.812		QuickStepEngine	Average Repetition Time 11.5221 s		
018-02-21 13:29:36.021		QuickStepEngine	starting Execution Phase: 'Testrum After'		
018-02-21 13:29:40.019		QuickStepEngine	Finished Execution Phase: 'Testrum After'		
018-02-21 13:29:40.019		QuickStepEngine	Total Test Time 17.3636 s		
1818-02-21 13:29:40.021	•	QuickStepEngine	Test finish		- 4
018-02-21 13:29:40.021		QuickStepEngine	Testrun finished @ 2018/02/21;13:29:40:019		
018-02-21 13:29:40.021	0	QuickStepEngine	Shutting down the Execution Environment and	SequencerPlease wait.	
(<u> </u>		
est Procedure Debugger					v 0
Remove all Breakpoints				(Help
Test Procedure Before	Test Proce	dure Test Procedur	e After	Test Procedure Browser	a 🐒
₩ ₽	() () () () () () () () () () () () () (Sacarban Sac	away Generator	Man Procedure Before (R) Testen Before (II) (R) DUT Loop Before (II) (R) Test Procedure (R) Test Procedure Before (II) (R) Test Procedure Before (R) DUT Loop After (II) (R) Testrum After (II) (R) Testrum After (II)	4 Properties

Signal analysis with R&S®QuickStep test executive software

Continuous average power measurements down to -130 dBm - precise and fast

Conventional diode power sensors reach their physical limits at approx. –70 dBm. Fast measurements degrade accuracy, especially at low power levels, since the noise content measured by these sensors is relatively high. As a result, either speed or accuracy has to be sacrificed.

The receiver based architecture of the R&S®NRQ6 eliminates this problem. This concept lowers measurement noise. In addition, the sensor's ability to perform band-limited measurements reduces the noise floor. These characteristics enable high-precision, high-speed measurements down to -130 dBm.

I/Q data capturing for RF vector signal analysis

The R&S®NRQ6 can be used as a standalone RF frontend to capture vector-modulated I/Q signals.

With the optional R&S[®]NRQ6-K1 I/Q data interface, captured I/Q data can be read out using SCPI commands. The data is demodulated and analyzed using external software, e.g. R&S[®]VSE.

With R&S[®]NRQ6 version 2.10 and R&S[®]VSE version 1.70, direct data aquisition and control is possible.

&S VSE Edit Input & Output Meas Setu		Parks and and and									- 0 ×
Edit input a Cutput Meas Setu			None T	XXXVV			or 🕨 🖾 🖉 🖓	ko 2			
asurement Group Setup Instrum		C SG NR 2	U Hone A J	()0)0			• • • • • • • • •	AL 1			
Surchast of oup octop	e x										
ew Group O Channel	Replace Channel	Ref Level 0.00 dBm Freq 1.0 GHz	Mode Downlin	k, 100 MHz Captu	re Time 20.1 ms BW	P/SS All					
	0.0										
🗹 Group 1 📃 🕨	- II @ O X	Att 0 dB	Frame Count	1 of 1(1) Frame	e 1						
		 5G NR 2: 1 Capture Buffer 				Older 5 k	* 5G NR 2: 3 EVM vs Can	ine			• 1 SS • 2 BWP0 🗗
🖲 🔄 SG NR 2 🛛 🚺 🕨	• II 🖒 🖲 🖂	Frame Start Offset : 2.341190585 ms					 BO INK 2: S EVM VS Call 	net .			0155 0 2 BWP0 BP1
•		20 0811									
		0 dBm					15 %				
		u2016im ether train did states and			a da la biene store da s		10.%				
		-40 dBm					10.55				
		160 deman and a dashed balance is to the	kontra antise to tasta disease a	a data bita a	and a store or	ois constation	5 %				
					والإذراق فيخت التلاز	بالالالالاتية فتت					
		-au aim -	2.01	mc/		20.1 ms	0.0.47		9.83 MHz/		98.28 MH
		 5G NR 2: 6 Constellation Diagram 	2.01	ing			 SG NR 2: 5 Power Spec 	tours	9.03 Millio		• 1 Cirw 🗗
		Points Measured : 915924						adm			
							-60 dBm/Hz				
							-80 dBm/Hz				
							-100 dBm/Hz				
							-Too delityHz				
							-120 dBm/Hz				
							0.0 Hz		12.29 MHz/		122.88 MH
		 5G NR 2: 2 Result Summary 			Collector -	1 Averaged 🗗 🐨		South all M. Casarina	12:29 MHg/		122.88 MF
		Frame Results Averaged	Mean	Limit	Max	Min	PSS	555	PBCH	PTRS	
		EVM PDSCH QPSK (%)		18.50	0.77	0.77	CORESET DMRS	PBCH DMRS	PDSCH DMRS	P D S C H	Not Used
		EVM PDSCH 16QAM (%) EVM PDSCH 64QAM (%)		13.50 9.00							
		EVM PDSCH 256QAM (%)		4.50							
		EVM All (%)	0.74		0.74	0.74					
		EVM Phys Channel (%) EVM Phys Signal (%)	0.75		0.43	0.75					
		Frequency Error (Hz)	139.66		139.66	139.66					
		Sampling Error (ppm) I/Q Offset (dB)	0.44		0.44	0.44					
		I/Q Gain Imbalance (dB)	-50.37		-50.37	-50.37					
		I/Q Quadrature Error (*)									
		OSTP (dBm) Power (dBm)	-29.98		-29.98	-29.98					
		Crest Factor (dB)	11.75		25.50	23.30					
ay 🖒 CONT	Record										

5G NR measurement with R&S®NRQ6 and R&S®VSE

Automated, cloud based data processing and analysis is also possible using the R&S®Quickstep test executive software to control any analysis tool in order to measure error vector magnitude (EVM), adjacent channel leakage ratio (ACLR) and other TX performance parameters.

Trace measurements

A detailed trace display is necessary for precise analysis of short pulses. With an inherent rise/fall time of 13 ns at a resolution bandwidth of 50 MHz, for example, the R&S®NRQ6 can easily measure steep-edged pulses.

Ultrafast triggered measurements


Triggered measurements in particular call for ever higher measurement speeds over an extended period of time. The R&S®NRQ6 contains a powerful FPGA and a large memory to meet these requirements. More than 100 000 triggered readings can be stored in a buffer in 200 ms – corresponding to a measurement speed of 500 000 readings/s – and transferred to a control PC.

Easy ACLR measurements

The frequency selective power sensor is perfect for adjacent channel leakage ratio (ACLR) measurements, which are frequently required in mobile communications. The ACLR measurement function is accessible from the web GUI and automatically sets one of the predefined 3GPP or LTE filters. The R&S®NRQ6 achieves an ACLR performance of typically –63 dBc for a 20 MHz LTE signal at –20 dBm.



Trace measurement on a pulsed signal



ACLR measurement on a 3.84 MHz 3GPP signal

EASY OPERATION

Intuitive web GUI

The R&S[®]NRQ6 is connected to the LAN via a PoE+ switch. The sensor includes an integrated web server. The intuitive web GUI can be operated from any web browser.

Diverse autoset functions

Diverse autoset functions are available to simplify configuration of the main measurement parameters. The measurement frequency and signal bandwidth are automatically determined and set. As a result, even unknown signals are detected and average power is measured accurately.

Depending on the input level, the 30 dB RF input attenuator is automatically switched on or off to configure the optimal power measurement range.

The trace mode also offers autoset functions. For example, the time scale (x-axis) and the power scale (y-axis) can be optimally configured. A trigger is set automatically, ensuring stable display of the measured signal.



Rear view of the R&S®NRQ6 with hardware interfaces

Automatic frequency tracking

A frequency tracker automatically sets the center frequency to facilitate measurements on narrowband signals with varying center frequency. This ensures that the measured signal is always within the selected measurement bandwidth.

Spectrum display for signal check

Since power measurements are performed only in the set frequency range (defined by the center frequency and bandwidth), the settings must be checked to make sure they are correct. The signal check function provides a graphical display of the measured signal's spectrum and the set bandwidth. Users can see at a glance if the measured signal is within the set frequency range.


Hardware interfaces for user convenience

The R&S[®]NRQ6 can be easily integrated into a test system. Remote operation is possible via LAN and USB.

The trigger I/O port can accept an external trigger signal or distribute an internally generated trigger signal to other R&S®NRQ6 power sensors.

An external LO signal can be fed to one of the R&S[®]NRQ6 power sensors, or the internal LO signal can be output and distributed to the other sensors.

The R&S®NRQ6 has a reference I/O port, e.g. for applying an external reference signal, and a sample clock I/O port.



Signal check for a 20 MHz LTE signal

APPLICATIONS

TX power calibration

To calibrate a DUT's transmit power, it is necessary to measure the frequency response at higher levels and the linearity down to minimum levels. The R&S®NRQ6 performs both measurements. The sensor not only stands out due to its high-precision power measurements, it also features excellent linearity of 0.02 dB. The R&S®NRQ6 is a compact, single-device solution for calibrating transmit power. No additional instruments or components such as a splitter and spectrum analyzer are needed. The sensor can be directly connected to the DUT; no cable is required. This solution provides better stability, lower mismatch and higher accuracy.

Band-limited power measurements on multistandard radios (MSR)


The R&S[®]NRQ6 can perform band-limited power measurements, i.e. it can measure the power on a selected transmission channel with a bandwidth up to 100 MHz, independent of neighboring channels. Band-limited power measurements can also be performed on base stations that support multiple mobile communications standards (MSR base stations), even if the user wants to measure only one standard.

Fast power servoing with R&S®NR06 and R&S®SGT100A


In component test applications, non-linear DUT characteristics make it difficult to set a specific output power. Conventional solutions based on SCPI via Ethernet are limited in speed. It takes several iterations to reach the desired power level, so the typical set time is 10 ms. This may be not adequate for production purposes.

The R&S[®]NRQ6-K2 power servoing option enables a dedicated serial communications channel between R&S[®]NRQ6 and R&S[®]SGT100A via the trigger connectors. Using this high-speed remote control directly between the FPGAs bypasses the speed limits of SCPI over Ethernet. It typically takes 1 ms to 1.5 ms to set the desired power on the R&S[®]SGT100A.

The R&S®NRQ6 can additionally perform power measurements without disconnecting, simply by remote control commands. For maximum performance and speed in power servoing applications, the setup can be extended by additional equipment. In that case, the R&S®NRQ6 is dedicated to fast power servoing and the measurements are performed, e.g. with a spectrum analyzer (SPA) or a second R&S®NRQ6.




Power servoing with R&S®NRQ6 and R&S®SGT100A




Single-channel power measurement on an MSR signal

Optional SPA for maximum performance and speed



Calibration of multiple active antenna modules for beamforming

The synchronous phase coherent measurements (R&S®NRQ6-K3 option) are based on a master/slave relationship between two or more R&S®NRQ6. The master forwards the local oscillator (LO) signal and its clock signal (CLK) to the slave(s) and triggers the slave(s). All measurement results are performed relative to the master. A calibration step before the first measurement helps eliminate the resulting group delay in the slave measurements. This is a simple and ideal solution for calibrating multiple active antenna modules for beamforming applications.




Results of an example multichannel phase difference measurement with one R&S®NRQ6 as master and eight R&S®NRQ6 as slaves. The master R&S®NRQ6 is only used for distribution of TRG, CLK and LO signals.

HIGH MEASUREMENT SPEED AT LOWEST SIGNAL LEVELS

Every diode based power meter works without band limitation in the specified frequency range. This feature makes the sensors very flexible, but has a significant impact: the noise contribution. Precise measurement of low-power signals requires averaging of multiple measurement samples and this increases the measurement time. The diode based power meter technology is mature, and therefore only limited product innovations are possible.

Precise and fast measurement of very low-power signals below –30 dBm was not possible until now. The receiver technology based R&S®NRQ6 has significantly increased the measurement speed for low-power signals. Compared with R&S[®]NRP8S, the R&S[®]NRQ6 reduces the measurement time by a factor of approx. 20000 for a CW signal with a power level of –50 dBm and noise contribution of 0.01 dB.

Comparison of measurement time for a CW signal



¹⁾ Measurement time: integration time + impulse response of the RF filter (B_{gF}).

²⁾ Lower measurement limit for RBW = 10 MHz.

³⁾ The discontinued R&S®NRP-Z11 is the direct predecessor of the R&S®NRP8S.

SPECIFICATIONS

Definitions


General

Product data applies under the following conditions:

- ▶ Three hours storage at ambient temperature followed by 30 minutes warm-up operation
- Specified environmental conditions met
- Recommended calibration interval adhered to
- ► All internal automatic adjustments performed, if applicable

Specifications with limits

Represent warranted product performance by means of a range of values for the specified parameter. These specifications are marked with limiting symbols such as $<, \leq, >, >, \geq, \pm$, or descriptions such as maximum, limit of, minimum. Compliance is ensured by testing or is derived from the design. Test limits are narrowed by guard bands to take into account measurement uncertainties, drift and aging, if applicable.



Specifications without limits

Represent warranted product performance for the specified parameter. These specifications are not specially marked and represent values with no or negligible deviations from the given value (e.g. dimensions or resolution of a setting parameter). Compliance is ensured by design.

Typical data (typ.)

Characterizes product performance by means of representative information for the given parameter. When marked with <, > or as a range, it represents the performance met by approximately 80% of the instruments at production time. Otherwise, it represents the mean value.

Nominal values (nom.)

Characterize product performance by means of a representative value for the given parameter (e.g. nominal impedance). In contrast to typical data, a statistical evaluation does not take place and the parameter is not tested during production.

Measured values (meas.)

Characterize expected product performance by means of measurement results gained from individual samples.

Uncertainties


Represent limits of measurement uncertainty for a given measurand. Uncertainty is defined with a coverage factor of 2 and has been calculated in line with the rules of the Guide to the Expression of Uncertainty in Measurement (GUM), taking into account environmental conditions, aging, wear and tear.

Device settings and GUI parameters are indicated as follows: "parameter: value".

Typical data as well as nominal and measured values are not warranted by Rohde&Schwarz.

In line with the 3GPP/3GPP2 standard, chip rates are specified in Mcps (million chips per second), whereas bit rates and symbol rates are specified in Mbps (million bits per second), kbps (thousand bits per second), Msps (million symbols per second) or ksps (thousand symbols per second), and sample rates are specified in Msample/s (million samples per second). Mcps, Mbps, Msps, ksps and Msample/s are not SI units.

R&S®NRQ6 power sensor signal flow from RF input connector to result processing




Specifications				
Frequency range		50 MHz to 6.0 GHz		
Impedance matching (SWR)	50 MHz to 100 MHz	< 1.20		
	> 100 MHz to 2.0 GHz	< 1.13		
	> 2.0 GHz to 6.0 GHz	< 1.20		
Power measurement range	dynamic range dependent on RBW (10 Hz to 400 MHz)	-130 dBm to +20 dBm		
Damage level	max. average power	1 W (+30 dBm) contin	uous	
	max. peak envelope power	2 W (+33 dBm) for ma		
	max. DC voltage	± 20 V	•	
RF input attenuation	6	0 dB, 30 dB		
Measurement subranges	RF input attenuation $= 0 \text{ dB}$	–130 dBm to –10 dBm		
	RF input attenuation = 30 dB	-100 dBm to +20 dBm		
Resolution bandwidth (RBW) ¹⁾	single-sideband (SSB) mode	10 Hz to 40 MHz		
	zero IF mode			
	(RF input frequency ≥ 400 MHz)	50 MHz, 80 MHz, 100	MHz, 400 MHz	
Acquisition	sample rate	119 MHz to 121 MHz ²)	
Displayed average noise level (DANL) 3)			
	RF input attenuation = 0 dB			
	50 MHz to 100 MHz	< –148 dBm (1 Hz)		
	> 100 MHz to 400 MHz	< –153 dBm (1 Hz)		
	> 400 MHz to 2.4 GHz	< –156 dBm (1 Hz)		
	> 2.4 GHz to 6.0 GHz	< –153 dBm (1 Hz)		
	RF input attenuation = 30 dB			
	50 MHz to 100 MHz	< -118 dBm (1 Hz)		
	> 100 MHz to 400 MHz	< –123 dBm (1 Hz)		
	> 400 MHz to 2.4 GHz	< -126 dBm (1 Hz)		
	> 2.4 GHz to 6.0 GHz	< –121 dBm (1 Hz)		
Uncertainty for absolute power meas	surements 4)			
	operating temperature range	+20°C to +25°C	+15°C to +35°C	0°C to +50°C
	RF input attenuation = 0 dB			
	50 MHz to 100 MHz	0.156 dB	0.167 dB	0.211 dB
	> 100 MHz to 400 MHz	0.130 dB	0.143 dB	0.192 dB
	> 400 MHz to 3 GHz	0.080 dB	0.100 dB	0.163 dB
	> 3 GHz to 6 GHz	0.092 dB	0.110 dB	0.169 dB
	RF input attenuation = 30 dB			
	50 MHz to 100 MHz	0.176 dB	0.189 dB	0.237 dB
	> 100 MHz to 400 MHz	0.147 dB	0.162 dB	0.216 dB
	> 400 MHz to 3 GHz	0.093 dB	0.114 dB	0.183 dB
	> 3 GHz to 6 GHz	0.105 dB	0.125 dB	0.190 dB
Uncertainty for relative power n	neasurements ⁵⁾ between any two power			
,	RF input attenuation = 0 dB			
	-60 dBm to -20 dBm	0.020 dB		
	RF input attenuation = 30 dB			
	-30 dBm to +10 dBm	0.020 dB		

Internal relevancy (in) security site 10 ⁴ Internal state frequency (in) RBW > 50 Mi/s 20 Mi /s 00 Mi /s ⁻¹ 0°C to +30°C 0°C to +50°C 0°C to +30°C 0°C to +50°C 0°C to +30°C 0°C to +50°C	Specifications					
InternetionerRRW a 40 Mebro200416-20 Mebro300 (F)If latesoperating temperator energie4.20°C to 4.25°C4.16°C to 4.35°C0.10°C to 4.50°CIs RRW a 40 Mebroby set 4.002.400 Mebroby set 4.002.400 Mebroby set 4.002.400 Mebro0.10°C to 4.50°CIs RRW a 40 Mebroby set 4.002.400 Mebroby set 4.003.400 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 40 Mebroby set 4.002.400 Mebroby set 4.003.400 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 400 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 400 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 400 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to 4.50°C0.10°C to 4.50°CIs RRW a 500 Mebroby set 4.000 Mebroby set 4.000 Mebro0.10°C to	-	accuracy	$\pm 1 \times 10^{-6}$			
IF fatnessoperating temperature range $\pm 20^\circ C to \pm 25^\circ C\pm 16^\circ C to \pm 35^\circ C0^\circ C to \pm 50^\circ CI RBW \pm 0.00 MizRBW \pm 0.00 Mizkp, \pm 20.02 dBkp, \epsilon \pm 0.03 dBkp, \epsilon \pm 0.08 dBI RBW \pm 0.00 Mizkp, \epsilon \pm 0.02 dBtpp, \epsilon \pm 0.03 dBkp, \epsilon \pm 0.08 dBI RBW \pm 0.00 Mizkp, \epsilon \pm 0.02 dBtpp, \epsilon \pm 0.03 dBtpp, \epsilon \pm 0.08 dBI RBW \pm 0.00 Mizkp, \epsilon \pm 0.02 dBtpp, \epsilon \pm 0.03 dBtpp, \epsilon \pm 0.08 dBI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.03 dBtpp, \epsilon \pm 0.08 dBI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI RBW \pm 0.00 Miztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI CHztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI CHztpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCtpp, \epsilon \pm 0.04 BCI CHz$		$RBW \le 40 MHz$	20 MHz to 30 MHz ⁶⁾			
• PF input frequency = 400 MHz Npx < 40.02 dB Npx < 40.03 dB Npx < 40.03 dB Npx < 40.03 dB Npx < 40.09 dB Image response operating temportation cange +20°C to +25°C +15°C to -35°C 0°C to +50°C 50 MHz to 100 MHz Vpc < -30 dBc Vpc < -30 dBc Vpc < -35 dBc Vpc < -35 dBc 50 MHz to 400 MHz Vpc < -40 dBc Vpc < -35 dBc Vpc < -35 dBc Vpc < -35 dBc 50 MHz to 400 MHz Vpc < -30 dBc Vpc < -40 dBc Vpc < -35 dBc Vpc < -35 dBc 50 MHz to 400 MHz Vpc < -30 dBc Vpc < -40 dBc Vpc < -40 dBc Vpc < -40 dBc 50 MHz to 61/2 Vpc < -40 dBc Vpc < -40 dBc Vpc < -40 dBc Vpc < -40 dBc 6 GHz Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc 1 GHz Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc 1 GHz Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc 1 GHz Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc Vpc < -70 dBc 1 GHz Vpc < -70 dBc		RBW ≥ 50 MHz	zero IF			
PRRW = 40 MHP NP, ≈ ±0.02 dB Np, ≈ ±0.03 dB Np, ≈ ±0.03 dB Image response operating temperature ange ±00°C to ±25°C ±15°C to ±35°C 0°C to ±50°C Image response operating temperature ange ±00°C to ±25°C ±15°C to ±35°C 0°C to ±50°C Image response ±00°MH² to 400 MH² Np, ≈ ±0.40 BC	IF flatness	operating temperature range	+20°C to +25°C	+15°C to +35°C	0°C to +50°C	
50 MHz to 100 MHz typ, <-30 dBc		 ► RBW ≤ 40 MHz ► RBW filter type: flat ► offset from center frequency 	typ. < ±0.02 dB	typ. < ±0.03 dB	typ. < ±0.08 dB	
> 100 MHz to 400 MHz typ. < -45 dBc	Image response	operating temperature range	+20°C to +25°C	+15°C to +35°C	0°C to +50°C	
b 400 MHz to 6 GHz typ, < -40 dBc t		50 MHz to 100 MHz	typ. < –30 dBc	typ. < -30 dBc	typ. < -25 dBc	
LQ phase noise at 1 kHz offset, measurement bardwith 1 Hz, measured at LO VO connector 400 MHz typ, < 98 dBc 400 MHz typ, < 92 dBc 2 CHz typ, < 90 dBc 4 CHz typ, < 86 dBc 4 CHz typ, < 86 dBc 4 CHz typ, < 86 dBc 4 CHz typ, < 74 dBc CO Leakage at RF input connector (I/O Frequency and frequencies of harmonics) FF input attenuation = 0 dB FF input attenuation = 0 dB Third-order intercept point (TO) '' FF input attenuation = 0 dB CO MHz typ, > +13 dBm A GHz typ, > +13 dBm A GHz typ, > +10 dBm A GHZ typ, >		> 100 MHz to 400 MHz	typ. < -45 dBc	typ. < -40 dBc	typ. < -35 dBc	
400 MHztyp. < -98 dBc1 GHztyp. < -98 dBc		> 400 MHz to 6 GHz	typ. < -50 dBc	typ. < -45 dBc	typ. < -40 dBc	
1 GHzipp. < -92 dBc2 GHzipp. < -92 dBc	LO phase noise	at 1 kHz offset, measurement bandw	vidth 1 Hz, measured at	LO I/O connector		
2GHzYp. < -86 dBc6GHzYp. < -74 dBc		400 MHz	typ. < -98 dBc			
vp.vp.<-80 dBc6 GH2typ.<-74 dBc		1 GHz	typ. < -92 dBc			
$ \begin{array}{c c c } & \begin{tabular}{l c c c c c } & \begin{tabular}{l c c c c c c c } & \begin{tabular}{l c c c c c c c c c c c c c c c c c c c$		2 GHz	typ. < -86 dBc			
ID leakage at RF input connector (LD frequency and frequencies of harmonic intercept point (PDI) RF input attenuation = 0 dB IF - 3 GH2 typ. < -55 dBm		4 GHz	typ. < -80 dBc			
RF input attenuation = 0 dB4 < 3 GHz		6 GHz	typ. < -74 dBc			
i i SGHz typ. < -55 dBm	LO leakage at RF input connector (LO frequency and frequencies of harr	nonics)			
3 GHz ≤ f ≤ 6 GHz typ. < -45 dBm		RF input attenuation = 0 dB				
RF input attenuation = 30 dB $f < 3$ GHz typ. < -75 dBm		f < 3 GHz	typ. < –55 dBm			
f < 3 GHz typ. < -75 dBm 3 GHz ≤ f ≤ 6 GHz typ. < -65 dBm		$3 \text{ GHz} \le f \le 6 \text{ GHz}$	typ. < –45 dBm			
No. No. No. Third-order intercept point (TOI) " RF input attenuation = 0 dB 400 MHz typ. > +13 dBm 400 MHz typ. > +13 dBm 400 MHz typ. > +12 dBm 2 GHz typ. > +10 dBm 2 GHz typ. > +10 dBm 6 GHz typ. > +5 dBm 6 GHz typ. > +5 dBm 6 GHz typ. > +5 dBm 6 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 6 GHz 10 GHz		RF input attenuation = 30 dB				
Third-order intercept point (TOI) ⁿ RF input attenuation = 0 dB 400 MHz typ. > +13 dBm 400 MHz typ. > +12 dBm 2 GHz typ. > +10 dBm 2 GHz typ. > +8 dBm 6 GHz typ. > +5 dBm 8 GHz typ. > +5 dBm 1 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 1 GHz typ. > +43 dBm 1 GHz typ. > +43 dBm 2 GHz typ. > +43 dBm 6 GHz typ. > +38 dBm 6 GHz typ. > +38 dBm 6 GHz typ. > +36 dBm 6 GHz typ. > +30 dBm 6 GHz typ. > +20 dBm 6 GHz typ. > +30 dBm		f < 3 GHz	typ. < –75 dBm			
RF input attenuation = 0 dB 400 MHz typ. > +13 dBm 1 GHz typ. > +12 dBm 2 GHz typ. > +10 dBm 4 GHz typ. > +10 dBm 6 GHz typ. > +5 dBm 6 GHz typ. > +5 dBm MHz typ. > +43 dBm 400 MHz typ. > +43 dBm 400 MHz typ. > +43 dBm 400 MHz typ. > +43 dBm 2 GHz typ. > +43 dBm 2 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 6 GHz typ. > +35 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHT) Filteruation = 0 dB Filteruation = 0 dB Filteruation = 0 dB GHz 1 GHz		$3 \text{ GHz} \le f \le 6 \text{ GHz}$	typ. < -65 dBm			
400 MHz typ. > +13 dBm 1 GHz typ. > +12 dBm 2 GHz typ. > +10 dBm 4 GHz typ. > +8 dBm 6 GHz typ. > +5 dBm 6 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 400 MHz typ. > +43 dBm 1 GHz typ. > +42 dBm 2 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm Second harmonic intercept point (SHI V Second harmonic intercept point (SHI Second harmonic intercept point (SHI Second harmonic intercept point (SHI	Third-order intercept point (TOI) 7)					
1 GHz ivp. > +12 dBm 2 GHz tvp. > +10 dBm 4 GHz tvp. > +8 dBm 6 GHz tvp. > +5 dBm 8 F input attenuation = 30 dB 400 MHz 4 00 MHz tvp. > +43 dBm 1 GHz tvp. > +43 dBm 2 GHz tvp. > +43 dBm 2 GHz tvp. > +43 dBm 2 GHz tvp. > +43 dBm 6 GHz tvp. > +35 dBm 6 GHz tvp. > +35 dBm Formutation = 0 dB 1 GHz tvp. > +45 dBm 2 GHz tvp. > +33 dBm 4 GHz tvp. > +33 dBm 6 GHz tvp. > +25 dBm 6 GHz tvp. > +25 dBm 6 GHz tvp. > +25 dBm 6 GHz tvp. > +70 dBm 6 GHz		RF input attenuation $= 0 \text{ dB}$				
2 GHz typ. > +10 dBm 4 GHz typ. > +8 dBm 6 GHz typ. > +5 dBm 6 GHz typ. > +5 dBm RF input attenuation = 30 dB 400 MHz 4 00 MHz typ. > +43 dBm 1 GHz typ. > +42 dBm 2 GHz typ. > +40 dBm 4 GHz typ. > +43 dBm 6 GHz typ. > +30 dBm 6 GHz typ. > +45 dBm 2 GHz typ. > +45 dBm 2 GHz typ. > +43 dBm 6 GHz typ. > +43 dBm 6 GHz typ. > +45 dBm 2 GHz typ. > +45 dBm 6 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm 6 GHz typ. > +25 dBm 6 GHz typ. > +25 dBm 6 GHz typ. > +70 dBm 6 GHz typ. > +63 dBm 4 GHz typ. > +63 dBm 4 GHz <td></td> <td>400 MHz</td> <td>typ. > +13 dBm</td> <td></td> <td></td>		400 MHz	typ. > +13 dBm			
4 GHz typ. > +8 dBm 6 GHz typ. > +5 dBm RF input attenuation = 30 dB 400 MHz 400 MHz typ. > +43 dBm 2 GHz typ. > +42 dBm 2 GHz typ. > +42 dBm 6 GHz typ. > +40 dBm 6 GHz typ. > +38 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) FF input attenuation = 0 dB RF input attenuation = 0 dB 2 GHz typ. > +45 dBm 2 GHz typ. > +45 dBm 6 GHz typ. > +38 dBm 6 GHz typ. > +38 dBm 8 F input attenuation = 0 dB 1 GHz 1 GHz typ. > +38 dBm 6 GHz typ. > +38 dBm 1 GHz typ. > +38 dBm 6 GHz typ. > +38 dBm 1 GHz typ. > +30 dBm 6 GHz typ. > +30 dBm 1 GHz typ. > +25 dBm 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +56 dBm 6 GHz typ. > +56 dBm			typ. > +12 dBm			
6 GHztyp. > +5 dBmRF input attenuation = 30 dB400 MHztyp. > +43 dBm1 GHztyp. > +42 dBm2 GHztyp. > +40 dBm6 GHztyp. > +38 dBm6 GHztyp. > +35 dBmSecond harmonic intercept point (SH)F input attenuation = 0 dB1 GHz1 GHztyp. > +45 dBm2 GHztyp. > +45 dBm2 GHztyp. > +45 dBm6 GHztyp. > +45 dBm6 GHztyp. > +25 dBm6 GHztyp. > +25 dBm6 GHztyp. > +25 dBm6 GHztyp. > +25 dBm6 GHztyp. > +26 dBm6 GHztyp. > +26 dBm6 GHztyp. > +26 dBm6 GHztyp. > +26 dBm6 GHztyp. > +60 dBm6 GHztyp. > +63 dBm6 GHztyp. > +56 dBm		2 GHz				
RF input attenuation = 30 dB 400 MHz typ. > +43 dBm 1 GHz typ. > +42 dBm 2 GHz typ. > +40 dBm 4 GHz typ. > +38 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) FF input attenuation = 0 dB FF input attenuation = 0 dB C GHz 1 GHz <t< td=""><td></td><td>4 GHz</td><td></td><td></td><td></td></t<>		4 GHz				
400 MHz typ. > +43 dBm 1 GHz typ. > +42 dBm 2 GHz typ. > +40 dBm 6 GHz typ. > +38 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) Friput attenuation = 0 dB Second harmonic intercept point (SHI) Friput attenuation = 0 dB Second harmonic intercept point (SHI) Second harmonic intercept point (SHI) <td colspan<="" td=""><td></td><td>6 GHz</td><td>typ. > +5 dBm</td><td></td><td></td></td>	<td></td> <td>6 GHz</td> <td>typ. > +5 dBm</td> <td></td> <td></td>		6 GHz	typ. > +5 dBm		
1 GHz typ. > +42 dBm 2 GHz typ. > +40 dBm 4 GHz typ. > +38 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) Friput attenuation = 0 dB RF input attenuation = 0 dB 1 GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 2 GHz typ. > +38 dBm 6 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB I GHz 1 GHz typ. > +70 dBm 1 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm						
2 GHz typ. > +40 dBm 4 GHz typ. > +38 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) F input attenuation = 0 dB RF input attenuation = 0 dB 2 GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 4 GHz typ. > +38 dBm 6 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB typ. > +25 dBm 1 GHz typ. > +70 dBm 1 GHz typ. > +70 dBm 4 GHz typ. > +63 dBm 4 GHz typ. > +63 dBm						
wn wn wn 6 GHz typ. > +38 dBm 6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) FF input attenuation = 0 dB RF input attenuation = 0 dB 1 GHz 2 GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 6 GHz typ. > +38 dBm 6 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB typ. > +25 dBm 1 GHz typ. > +70 dBm 2 GHz typ. > +70 dBm 4 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm						
6 GHz typ. > +35 dBm Second harmonic intercept point (SHI) RF input attenuation = 0 dB 1 GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 4 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB typ. > +25 dBm RF input attenuation = 30 dB typ. > +70 dBm 1 GHz typ. > +70 dBm 4 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm			,,			
Second harmonic intercept point (SHI) RF input attenuation = 0 dB I GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 4 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB 1 GHz typ. > +70 dBm 1 GHz typ. > +70 dBm 4 GHz typ. > +70 dBm 4 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm						
RF input attenuation = 0 dB 1 GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 4 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB I GHz 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm			typ. > +35 dBm			
1 GHz typ. > +45 dBm 2 GHz typ. > +38 dBm 4 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm	Second harmonic intercept point (S					
2 GHz typ. > +38 dBm 4 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm						
4 GHz typ. > +30 dBm 6 GHz typ. > +25 dBm RF input attenuation = 30 dB 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm						
6 GHz typ. > +25 dBm RF input attenuation = 30 dB 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm			,1			
RF input attenuation = 30 dB 1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm			,1			
1 GHz typ. > +70 dBm 2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm			typ. > +25 dBm			
2 GHz typ. > +63 dBm 4 GHz typ. > +55 dBm						
4 GHz typ. > +55 dBm						
6 GHz typ. > +50 dBm			,,			
		6 GHz	typ. > +50 dBm			

Other characteristics				
Measurand		power of incident wave		
		power of source (DUT) into 50 $\boldsymbol{\Omega}$		
RF input connector		N (male)		
Measurement functions		 continuous average trace adjacent channel leakage ratio (A I/Q trace 	ACLR)	
Continuous average function	measurand	average power over acquisition inter	rval	
	aperture	8.3 ns to 30 s (depending on RBW)		
	duty cycle correction ⁸⁾	0.001% to 100.0%		
	capacity of measurement buffer ⁹⁾	1 reading to 8192 readings		
Trace function	measurand	average power over pixel(s)		
	acquisition			
	length	8.3 ns to 30 s (depending on RBW)		
	start (referenced to delayed trigger)	-15.0 s to $+15.0$ s (depending on RE	3VV)	
	result			
	number of pixels	1 to 8192		
	resolution	\geq 8.3 ns (sample period depending of	on RBW)	
Adjacent channel leakage ratio (ACLR)	measurand	power ratio		
	supported standards	 3GPP (3.84 MHz) EUTRA/LTE (5 MHz, 10 MHz, 15 	MHz, 20 MHz)	
	acquisition length	1 ms to 40 ms		
	dynamic range	test model	level = –20 dBm, carrier frequency = 2 GHz	
		3GPP FDD, test model 1, 64 DPCH	–69 dBc (meas.) ¹⁰⁾	
		EUTRA/LTE 5 MHz	–68 dBc (meas.) ¹⁰⁾	
		EUTRA/LTE 10 MHz	–65 dBc (meas.) ¹⁰⁾	
		EUTRA/LTE 20 MHz	–63 dBc (meas.) ¹⁰⁾	
Phase coherent measurements	measurand	standard deviation of 10 consecutive slave R&S®NRQ6	e phase differences between two	
	supported standards	5G NR, LTE, MCCW, CW		
		MCCW signal level	temperature: +20 °C signal bandwidth: 100 MHz number of carriers: 100 carrier spacing: 1 MHz center frequency: 3.5 GHz	
		–50 dBm	< 0.05° (meas.)	
		–60 dBm	< 0.08° (meas.)	
		–70 dBm	< 0.16° (meas.)	
		, o dBill		

R&S®NR06 ALCR performance at 2 GHz over power; EUTRA/LTE 20 MHz (meas., noise correction on)



Other characteristics		
I/Q trace function	measurand	I/Q complex voltage
	prerequisite	R&S®NRQ6-K1 option
	acquisition length	8.3 ns to 30 s (depending on RBW)
	result	
	output sample rate	100 Hz to 120 MHz (continuously variable, impacts effective RBW)
	number of I/Q pairs	1 to 15000000
Power servoing with R&S®SGT100A	tolerance	0.01 dB to 3 dB
R&S®SGT100A firmware ≥4.65	sensor aperture time	10 µs to 100 ms
	setting time	typ. < 1.5 ms (tolerance = 0.1 dB, aperture = 100 μ s, remote control via instrument driver)
	level tracking interval	500 ms
Triggering	supported measurement functions	continuous average, trace, I/Q trace
	source	 INTernal: internal test signal EXTernal2: coaxial trigger I/O (SMA (f) jack) EXTernal[1]: host interface trigger signal (8-pin male M12 connector) BUS: remote control event (*trg)
	dropout	0 s to 10 s (depending on RBW)
	slope (external, internal)	positive/negative
	delay	–5 s to +10 s ¹¹⁾ (depending on RBW)
	hold-off	0 s to 10 s (depending on RBW)
	resolution (delay, hold-off, dropout)	≥ 8.3 ns (depending on RBW)
	INTernal trigger threshold level	
	range	–110 dBm to +20 dBm
	accuracy	identical to uncertainty for absolute power measurements
	hysteresis	0 dB to 10 dB
	trigger jitter	 internal trigger: ≥ 8.3 ns (depending on RBW) external trigger: 8.3 ns
Averaging filter	parameters	
	supported measurement functions	continuous average, trace
	averaging count	1 to 65536
	result output	
	moving mode	continuous result output, independent of averaging count
	repeat mode	final result only
RF input attenuation correction	function	corrects the measurement result using a fixed factor (dB offset)
	range	-200 dB to +200 dB
	-ZKU interface cable (requires additiona	l PoE+ power supply at LAN interface) B sensor hub (requires additional PoE+ power supply at LAN interface)
	mechanical	8-pin male M12 connector (A-coded)
	power supply	+5 V/0.1 A (USB low-power device; requires additional PoE+ power supply)
	speed	high-speed and full-speed mode in line with USB specification
	remote control protocols	USB test and measurement class (USBTMC)
	trigger input EXTernal[1]	differential (0 V/+3.3 V)
	reference clock	
	signal level	LVDS
	input frequency	20 MHz
	permissible total cable length	≤ 5 m
Ethernet interface (LAN PoE+)	mechanical	RJ-45 jack
	power supply	power over Ethernet (PoE+) class 4
	speed	10/100/1000 Mbit/s
	remote control protocols	VXI-11, HiSLIP (high-speed LAN instrument protocol), SCPI-RAW (port 5025)
	permissible cable length	≤ 100 m
	. 0	

Other characteristics		
Trigger 2 I/O (TRIG2)	mechanical	SMA (f) jack
	impedance	
	input	10 k Ω or 50 Ω (software controlled)
	output	50 Ω
	signal level	
	input	compatible with 3 V or 5 V logic, max. –1 V to +6 V
	output	\geq 2 V into 50 Ω load, max. 5.3 V
Reference I/O (REF)	mechanical	SMA (f) jack
	impedance	
	input/output	50 Ω
	signal level	
	input	≥ -10 dBm
	output	≥ +7 dBm
	frequency	
	input	10 MHz
	output	10 MHz
Clock I/O (CLK)	mechanical	SMA (f) jack
	impedance	
	input/output	50 Ω
	signal level	
	output	≥ -10 dBm
	frequency	
	output	119 MHz to 121 MHz
Local oscillator I/O (LO)	mechanical	SMA (f) jack
	impedance	
	input/output	50 Ω
	signal level	
	input	≥ -5 dBm
	output	≥ 0 dBm
	frequency	
	input/output	70 MHz to 6.03 GHz

General data		
Temperature ranges ¹²⁾	operating temperature range	0°C to +50°C
	storage temperature range	-40°C to +85°C
Climatic resistance	damp heat	+25°C/+55°C cyclic at 95% relative humidity with restrictions: noncondensing, in line with EN 60068-2-30
Mechanical resistance	vibration	
	sinusoidal	5 Hz to 55 Hz, 0.15 mm amplitude, 1.8 g at 55 Hz, 55 Hz to 150 Hz, 0.5 g constant, in line with EN 60068-2-6
	random	8 Hz to 650 Hz, 1.9 g (RMS), in line with EN 60068-2-64
	shock	45 Hz to 2 kHz, max. 40 g shock spectrum, in line with MIL-STD-810E, method 516.4, procedure I
Air pressure	operating	795 hPa (2000 m) to 1060 hPa
	transport	566 hPa (4500 m) to 1060 hPa
Electromagnetic compatibility		 harmonized standards complied with: EN 61326-1 EN 61326-2-1 EN 55011 (class B)
Calibration interval	recommended	2 years
Dimensions	$W \times H \times D$	98 mm × 47 mm × 146 mm (3.85 in × 1.85 in × 5.75 in)
Weight		0.50 kg (1.10 lb)

 $^{1\!/}$ By default, the discrete RBW filter selection mode (1, 2, 3, 5, \dots steps) is active.

Alternatively, steplessly variable RBW can be activated for RBW < 20 MHz. The SNR in the variable RBW selection mode may be slightly lower than in the discrete RBW selection mode.

²⁾ The sample rate is selected automatically.

³⁾ This applies to resolution bandwidths ≤ 300 kHz. For larger RBWs, spurious frequency response contributions might violate these limits at certain RF input frequencies. In addition, RBWs using zero IF mode might violate these limits due to DC offset contributions.

⁴⁾ Expanded uncertainty (k = 2) for absolute continuous average power measurements on CW signals centered within RBWs ≤ 10 MHz. Specifications include calibration uncertainty, aging, linearity and temperature effect.

Measurement noise must additionally be taken into account when measuring low powers. The contribution of measurement noise can be neglected below a two-sigma value of 0.01 dB.

For signal power levels less than 30 dB above the DANL at the selected RBW, a measurement bias must additionally be taken into account.

For power levels above +15 dBm/–15 dBm for 30 dB/0 dB RF input attenuation, respectively, uncertainty contributions due to intermodulation and other nonlinear effects must additionally be taken into account.

For RBWs below 1 kHz, LO phase noise contributions must additionally be taken into account.

For RBWs above 300 kHz, uncertainty contributions due to coherent spurious response frequencies (e.g. when a locked reference frequency is used) must additionally be taken into account.

When configuring the power sensor for use with an externally supplied LO signal, uncertainty contributions due to the signal integrity of the external LO signal must additionally be taken into account. Signal integrity includes properties such as frequency accuracy, and amplitude and phase stability.

Expanded uncertainty (k = 2) for relative power measurements on CW signals of identical frequency in continuous average mode for RBW ≤ 10 MHz. Specifications include aging and temperature effect.

Measurement noise must additionally be taken into account.

For signal power levels less than 30 dB above the DANL at the selected RBW, a measurement bias must additionally be taken into account.

For RBWs below 1 kHz, LO phase noise contributions must additionally be taken into account.

For RBWs above 300 kHz, uncertainty contributions due to coherent spurious response frequencies (e.g. when a locked reference frequency is used) must additionally be taken into account.

When configuring the power sensor for use with an externally supplied LO signal, uncertainty contributions due to the signal integrity of the external LO signal must additionally be taken into account. Signal integrity includes properties such as frequency accuracy, and amplitude and phase stability.

⁶⁾ The IF frequency is selected automatically.

- ⁷⁾ Measurements were performed using two-tone signals separated by 2 MHz.
- ⁸⁾ For average power measurements of periodic bursts.
- ⁹⁾ To increase measurement speed, the power sensor can be operated in buffered mode. In this mode, measurement results are stored in a buffer of user-definable size and output as a data block when the buffer is full. For further information, see application sheet "R&S®NRQ6. Fast Pulse Power Measurement" (1178824202).

¹⁰⁾ Noise correction improves the ACLR by typ. 5 dB, depending on the level and bandwidth.

¹¹⁾ In I/Q trace mode, only positive trigger delays are supported.

¹²⁾ The operating temperature range defines the span of ambient temperature in which the instrument complies with specifications.

ORDERING INFORMATION

Designation	Туре	Order No.
Frequency selective power sensor	R&S®NRQ6	1421.3509.02
I/Q data interface	R&S®NRQ6-K1	1421.4705.02
Power servoing	R&S®NRQ6-K2	1421.4740.02
Phase coherent measurements	R&S®NRQ6-K3	1421.4770.02
Accessories		
Ten-port PoE+ switch	R&S®NRP-ZAP2	3639.1902.02
USB interface cable, length: 0.75 m	R&S®NRP-ZKU	1419.0658.02
USB interface cable, length: 1.50 m	R&S®NRP-ZKU	1419.0658.03
USB interface cable, length: 3.00 m	R&S®NRP-ZKU	1419.0658.04
USB interface cable, length: 5.00 m	R&S®NRP-ZKU	1419.0658.05
Six-pole interface cable, length: 1.50 m	R&S®NRP-ZK6	1419.0664.02
Six-pole interface cable, length: 3.00 m	R&S®NRP-ZK6	1419.0664.03
Six-pole interface cable, length: 5.00 m	R&S®NRP-ZK6	1419.0664.04
USB sensor hub	R&S®NRP-Z5	1146.7740.02
Documentation		
Documentation of calibration values (DCV)	R&S®DCV-1	0240.2187.06
Hardcopy of DCV (in combination with R&S®DCV-1 only)	R&S®DCV-ZP	1173.6506.02
Accredited calibration for R&S®NRQ6 power sensor	R&S®NRP-ACA	1419.0812.00

Warranty		
Base unit		3 years
R&S®NRP-ZAP2		5 years
All other items ¹⁾		1 year
Options		
Extended warranty, one year	R&S®WE1	
Extended warranty, two years	R&S®WE2	
Extended warranty with calibration coverage, one year	R&S [®] CW1	Please contact your local
Extended warranty with calibration coverage, two years	Rohde&Schwarz sales office.	
Extended warranty with accredited calibration coverage, one year		
Extended warranty with accredited calibration coverage, two years	R&S®AW2	

Extended warranty with a term of one and two years (WE1 and WE2)

Repairs carried out during the contract term are free of charge²). Necessary calibration and adjustments carried out during repairs are also covered.

Extended warranty with calibration coverage (CW1 and CW2)

Enhance your extended warranty by adding calibration coverage at a package price. This package ensures that your Rohde&Schwarz product is regularly calibrated, inspected and maintained during the term of the contract. It includes all repairs²⁾ and calibration at the recommended intervals as well as any calibration carried out during repairs or option upgrades.

Extended warranty with accredited calibration (AW1 and AW2)

Enhance your extended warranty by adding accredited calibration coverage at a package price. This package ensures that your Rohde&Schwarz product is regularly calibrated under accreditation, inspected and maintained during the term of the contract. It includes all repairs²⁾ and accredited calibration at the recommended intervals as well as any accredited calibration carried out during repairs or option upgrades.

- ¹⁾ For options that are installed, the remaining base unit warranty applies if longer than 1 year. Exception: all batteries have a 1 year warranty.
- ²⁾ Excluding defects caused by incorrect operation or handling and force majeure. Wear-and-tear parts are not included.

Service that adds value

- ► Worldwide
- ► Local und personalized
- ► Customized and flexibel
- Uncompromising quality
- ► Long-term dependability

Rohde & Schwarz

The Rohde&Schwarz electronics group offers innovative solutions in the following business fields: test and measurement, broadcast and media, secure communications, cybersecurity, monitoring and network testing. Founded more than 80 years ago, the independent company which is headquartered in Munich, Germany, has an extensive sales and service network with locations in more than 70 countries.

www.rohde-schwarz.com

Sustainable product design

- ► Environmental compatibility and eco-footprint
- ► Energy efficiency and low emissions
- Longevity and optimized total cost of ownership



Rohde & Schwarz training

www.training.rohde-schwarz.com

Regional contact

- Europe, Africa, Middle East | +49 89 4129 12345 customersupport@rohde-schwarz.com
- North America | 1 888 TEST RSA (1 888 837 87 72) customer.support@rsa.rohde-schwarz.com
- Latin America | +1 410 910 79 88 customersupport.la@rohde-schwarz.com
- Asia Pacific | +65 65 13 04 88 customersupport.asia@rohde-schwarz.com
- China | +86 800 810 82 28 | +86 400 650 58 96 customersupport.china@rohde-schwarz.com

R&S° is a registered trademark of Rohde & Schwarz GmbH & Co. KG Trade names are trademarks of the owners PD 3607.1888.12 | Version 04.01 | October 2019 (sk) R&S°NR06 Frequency Selective Power Sensor Data without tolerance limits is not binding | Subject to change © 2017 - 2019 Rohde & Schwarz GmbH & Co. KG | 81671 Munich, Germany 6607.1888.12 04.01 PDP 1 en